

College of Physicians & Surgeons of Alberta

Laser Equipment Companion Manual

September 2017



College of
Physicians
& Surgeons
of A l b e r t a

Owner Responsibilities

- **Radiation equipment cannot be used until a CPSA registration certificate has been issued.**
- Register equipment according to CPSA step-by-step instructions.
- Recertify equipment according to CPSA step-by-step instructions.
- Ensure a copy of the original CPSA *Certificate of Registration* for equipment is on-site at the facility w it is registered.
- Pay the one-time registration fee(s) and subsequent annual renewal fee(s) in a timely manner.
- Submit annual confirmation of laser equipment profile(s) with/without changes back to CPSA.
- Notify CPSA of equipment that is discarded, stored, sold, leased, or transferred* within 30 days:
 - Complete and submit Notification of Change form with section “C” completed
 - Return original CPSA *Certificate of Registration*
 - If required, advise CPSA of new owner name and/or new location
- If equipment is sold, *your* responsibility is to inform the new owner to register with CPSA prior to operating.*
- Inform CPSA prior to resuming use of stored equipment.
- Stored equipment must be re-tested by an Authorized Radiation Protection Agency (ARPA) and new CPSA certificate issued when:
 - Certificate’s expiry date has lapsed
 - OR**
 - Placed in storage for longer than **one month**
- Notify CPSA of relocated equipment - Complete and submit Notification of Change form with section “B” completed
- Notify CPSA of equipment modifications - Complete and submit Notification of Change form with section “A” completed
 - Re-testing may be required

*Radiation Protection Regulation 13.1 & 13.2

FAQ's

- **Why do class 3b/4 lasers require registration?**
 - Class 3b/4 lasers are classified as designated radiation equipment, as per Section 8 of the Radiation Protection Regulation.
- **Why does CPSA register laser equipment?**
 - We are a Radiation Health Administrative Organization (RHAO) in contract with Alberta Labour to administer the Radiation Equipment Program which is responsible for registering physician owned/operated x-ray equipment and class 3b/4 lasers.
- **If a laser is owned by the facility and used by a CPSA registered physician, is it registered at CPSA?**
 - Yes.
- **If a laser is owned by the facility and used by other personnel (i.e. estheticians, nurses) but not a physician, is it registered at CPSA?**
 - No, if the facility is not physician-owned it is registered with the Alberta Association for Safety Partnerships (AASP).
- **If a CPSA physician only uses the laser occasionally, is registration still required?**
 - Yes.
- **What is an ARPA?**
 - Authorized Radiation Protection Agency – facility and equipment inspectors.
 - ARPA's are a completely independent, **separate regulatory-identified group**.
 - ARPA's are not employed by the CPSA.
- **What happens if an ARPA deems our facility and/or equipment non-compliant?**
 - The non-compliance report is submitted to CPSA.
 - CPSA sends a follow-up letter to owner indicating registration/recertification process **will not** continue and equipment **must not** be used until:
 - Non-compliance issues resolved and ARPA may need to re-test for compliance
 - Report submitted **and**
 - CPSA registration certificate issued
 - **It is the sole responsibility of the owner to work with the ARPA to achieve facility and/or equipment compliance.**
 - Once compliance is achieved the ARPA will notify the CPSA; the registration/recertification process will continue.
- **Can I submit CPSA equipment registration applications to the ARPA?**
 - No.

- **Compliance verification testing is required when:**
 - A new laser is purchased.
 - Existing equipment or the facility's structure is modified.
 - Existing equipment is relocated within the same facility.
 - Existing equipment is transferred to another facility.

- **How do I determine if a laser is class 3b/4?**
 - Read the manufacturer's labels on equipment, or read owner manual.

- **Can individuals other than physicians use lasers for hair removal?**
 - Yes. The registered owner is responsible for ensuring all personnel who use the laser are appropriately trained as per manufacturer/vendor instructions/requirements.

- **Why can't I register laser equipment using the paper-based verification form?**
 - Only lasers in a recertification cycle can take advantage of the paper-based compliance verification process.

- **Why can't equipment be operated without a registration certificate?**
 - Operation without a valid CPSA registration certificate is in direct contravention of the *Radiation Protection Act*.
 - A stop-use directive will be issued; a fine may be levied on prosecution from Alberta Labour.

- **Do temporary/loaner lasers require registration?**
 - Yes; restrictions are placed on the CPSA registration certificate.

- **How do I maintain laser equipment registration?**
 - Registered owners receive profiles of all active equipment from CPSA.
 - Owners are required to review/update profiles – with or without changes – to CPSA.
 - Reviewing equipment profiles is **not the same process** as recertifying equipment.

- **What happens if equipment is not recertified by the expiry date on the certificate?**
 - Certificate expiry is in contravention of the *Radiation Protection Act*; a stop-use directive is issued and remains in effect until recertified by the CPSA.

- **How do I know when equipment is expiring?**
 - Certificates have an expiry date located on the bottom of the left side.
 - It is the **owner's responsibility** to ensure equipment is re-tested, recertified and new certificate issued prior to the expiry date.

- **What is the difference between a mobile and a portable laser?**
 - There is no difference.

- **Can I register one laser to more than one facility?**
 - Yes.

For additional information on:

- Registration
- Recertification
- Equipment modification/relocation
- Decommissioned equipment
- ARPA listing
- Fees

Please refer to the CPSA Class 3b/4 Laser Equipment web page - www.cpsa.ca/accreditation/laser-equipment